



التاريخ : 24/07/2024

## طرح عطاء رقم RFQ-H-11709

السادة / الشركات المختصة الراغبة بالمشاركة بالعطاء - المحترمين

الموضوع : Infusion Pump , حسب المواصفات .

بالإشارة إلى الموضوع أعلاه, يرجى الاطلاع على المواصفات المرفقة مع مراعاة الشروط التالية :-

1. تكتب الاسعار بالشيكلي الإسرائيلي وبشكل واضح في المكان المخصص حسب الجدول المرفق و تشمل قيمة الضريبة المضافة, في حال وجود تعارض بين سعر الوحدة ومجموع سعر الوحدة المشتق, فإن سعر الوحدة هو الذي سيعتمد .
2. يتم تقديم شهادة خصم مصدر في حالة تجاوز المبلغ قيمة 2500 شيكل أو ما يعادلها .
3. غرامة تأخير 2% عن كل اسبوع وتبدأ من اليوم الذي يلي موعد التسليم المحدد من قبلكم , و يحق للمستشفى إلغاء أمر الشراء والاحالة على مورد آخر دون تحمل اية مسؤولية تذكر في حالة تأخير التوريد ما يزيد عن اسبوعين .
4. تحديد فترة التوريد في عرض سعركم وعدم ذلك يعني أن التوريد فوري وايضاً أن تكون صلاحية العرض المقدم من طرفكم مدة لا تقل عن ثلاثة أشهر .
5. تسلم عروض الاسعار بالظرف المغلق فقط , على أن يكون عرض السعر يحوي فصلاً بين العرض المالي والعرض الفني أي أن العرض المالي منفصل عن العرض الفني .
6. ضرورة الرد على العطاء سلباً أو ايجاباً .
7. الدفع بعد 90 يوم من تاريخ الإستلام والفحص ومطابقة المواصفات الفنية المطلوبة.
8. يحق للمستشفى تجزئة العطاء وهي غير ملزمة بأقل الاسعار .
9. يحق للمستشفى إلغاء العطاء دون ابداء السبب ودون تحمل اية تكاليف مع إعادة رسوم نسخة العطاء للمشاركين في العطاء ان وجدت, لا تتحمل المستشفى تكاليف النقل والتحميل والتنزيل .
10. ضرورة ارفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر .
11. اخر موعد لتقديم عرض السعر هو يوم الثلاثاء الموافق 2024/07/30 الساعة العاشرة صباحاً (10:00am).
12. لاية إستفسارات متعلقة بالعطاء المذكور اعلاه , يرجى الاتصال بالسيد مدير دائرة اللوازم والمشتريات هاتف رقم (0097092389687) داخلي (6400) أو من خلال البريد الإلكتروني tender3@najah.edu فقط.

مع فائق الاحترام ,,,,

مدير دائرة اللوازم والمشتريات

إياد مكاي



التاريخ : 24/07/2024

## طرح عطاء رقم RFQ-H-11709

No	Item Code	Item	Unit	QTY	Brand name + code	Unit Price (Nis)	Total Price (Nis)
1	G-A- MDD00016	Infusion Pump حسب المرفق	Pcs	15			
Total Price (NIS)							

- يرجى من حضرتكم تحديد مدة التوريد
- ضرورة الالتزام بالتسعير على النموذج المرفق بالعطاء، والالتزام بالعملة المحددة (شيك).

Infusion pump, Qty: 15	Compliance Yes / No	Deviation
Good brand name with FDA or CE preferable (Europe, USA, Japan)		
Should have CB test and EMC test (60601.1 and 60601.2)		
Manufacturer:		
Source and origin:		
Model:		
<b>Infusion pump Specification:</b>		
State of the art technology and the latest version from the manufacturer and the latest and best technology on the market		
Suitable for micro / adult patients		
Microprocessor Controlled		
Heavy duty with Large LCD touch screen		
Please specify the unit weight?		
Should compatible with most of the IV set and Programmable to any IV tubing (Open system For IV set)		
Accuracy $\pm 4\%$ (with using all brands of IV set)		
Operation modes included: Rate, Time, Weight, Sequence, Trapezia, Loading dose, Drop, Micro modes.		
Flow rate: 0.01-2000ml/h		
Volume infused: 0.01-9999.99 ml		
Please specify the KVO rate?		
Occlusion level: minimum 3 levels		
Ingress protection (IP) must be high. Please specify your device's IP rating?		
The equipment should have the ability to purge air to remove any outgassed air		
Audible and Visual alarms		
The system has an alarm when EOI (endo of infusion) has reached		
Should has bubble detector to detect air in line		
Free flow protection		
With upstream and downstream pressure sensor		
Should have bolus infusion function		
Should have a drug library (drug list) and a safety drug use alarm if a drug dose error occurs		
The units should be able to be combined together		
The equipment should have an internal rechargeable battery backup for minimum 7 hours		
The power cable should be an open system (universal)		
It can operate with and without drop sensor		
Should have a handle for portability and a clamp for IV stand attachment		
AC input: 220V, 50Hz		
Complete with all accessories and consumables needed to work completely as specified		
<b>Training included in the price:</b>		
Operation training for the oncology nursing team must be provided on-site		
<b>Optional to be priced separately:</b>		
List price all accessories necessary to operate the solution as per the required specs listed above		
List all other accessories and supplies that you may recommend to NNUH and are considered optional items to be decided on purchasing at the discretion of NNUH		
Drop sensor		
Screen		
Handel		
Battery		
Docking station for 4 pumps		
Pump door		
<b>Your offer should contain the following:</b>		
1) Preventive maintenance schedule		
2) Reference list for the same model, the supplier must have 3 reference at least on local market or regional, and must prepare site visit for our team to evaluate the product.		
3) Compliance sheet		
4) Catalogues		
5) Data sheet		
6) Authorization letter		
7) CE or FDA and ISO certificates		
8) CB test and EMC test certificate		
9) The service and operational manuals must delivered as soft copy with the units		
10) Guarantee for 10 years for spare parts supply		
Warranty: 3 years, with commitment to conduct preventive maintenance (labor and parts) during the warranty period and according to manufacturer recommendations		