

التاريخ : 05/02/2025 م

تعديل كراسة طرح عطاء رقم – RFQ-H-11781

السادة / الشركات المختصة الراغبة بالمشاركة بالعطاء – المحترمين .

الموضوع : Ventilators ، حسب المواصفات المرفقة .

بالإشارة إلى الموضوع أعلاه، يرجى الاطلاع على جداول المواصفات المرفقة مع مراعاة الشروط التالية :-

1. تكتب الاسعار بالشيكال الإسرائيلي وبشكل واضح في المكان المخصص حسب الجدول المرفق وتشمل قيمة الضريبة المضافة، في حال وجود تعارض بين سعر الوحدة ومجموع سعر الوحدة المشتق، فإن سعر الوحدة هو الذي سيعتمد .
2. ثمن نسخة العطاء (250) شيكال إسرائيلي غير مستردة تدفع في الدائرة المالية مبنى الإدارة – الطابق الأول .
3. كفالة دخول عطاء بقيمة (1000) ألف دولار امريكي او ما يعادلها سارية المفعول لمدة ثلاثة شهور تأميناً نقدياً او كفالة بنكية او شيكاً مصدقاً وفي حالة عدم تقديم المطلوب لا ينظر في عرض سعركم .
4. كفالة حسن تنفيذ بقيمة 10% من قيمة العطاء في حالة الترسية تأميناً نقدياً او كفالة بنكية او شيكاً مصدقاً سارية المفعول لمدة ثلاثة اشهر.
5. رسوم الاعلان على من يرسو عليه العطاء .
6. ضرورة أن تكون الأجهزة المعروضة جديدة وليست مجددة .
7. ضرورة الالتزام بالتسعير على النموذج المرفق بالعطاء .
8. يتم تقديم شهادة خصم مصدر في حالة تجاوز المبلغ قيمة 2500 شيكال او ما يعادلها.
9. غرامة تأخير 2% عن كل اسبوع وتبدأ من اليوم الذي يلي موعد التسليم المحدد من قبلكم، وفي حال تأخير التوريد مدة تزيد عن اسبوعين يحق للمستشفى إلغاء أمر الشراء والاحالة على مورد آخر مع تحميلكم فروق الأسعار .
10. الدفع خلال 90 يوماً من تاريخ الاستلام والفحص ومطابقة المواصفات الفنية المطلوبة .
11. تحديد فترة التسليم في عرض سعركم وايضاً فترة الضمان وأن تكون صلاحية العرض المقدم من طرفكم مدة لا تقل عن ثلاثة أشهر .
12. تسلم عروض الاسعار في مكتب مدير دائرة اللوازم والمشتريات رقم (6400) مبنى الإدارة في مستشفى النجاح الوطني الجامعي بالظرف المختوم فقط وتوضع في صندوق العطاءات ، على أن يكون عرض السعر يحوي فصلاً بين العرض المالي والعرض الفني أي أن العرض المالي منفصل عن العرض الفني مع ضرورة كتابة رقم العطاء على المغلفين وتوضيح اي منهما العرض الفني والعرض المالي.
13. على كل من يرغب بالمشاركة بتقديم عرض سعر تقديم ما لديه من مواد او اجهزة او ماركات لما جاء اعلاه او البديل له والرد على العطاء سلباً او ايجاباً .
14. يحق للمستشفى تجزئة العطاء وهي غير ملزمة بأقل الأسعار .
15. يحق للمستشفى إلغاء العطاء دون ابداء السبب ودون تحمل اية تكاليف مع اعادة رسوم نسخة العطاء للمشاركين في العطاء ان وجدت، ولا تتحمل المستشفى تكاليف النقل والتحميل والتتزيل .
16. ضرورة ارفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر .
17. آخر موعد لتقديم عرض السعر هو يوم الأحد الموافق 2025-02-16 م الساعة الثالثة مساءً 15:00 pm .
18. لأية إستفسارات متعلقة بالعطاء المذكور اعلاه ، يرجى الاتصال بالسيد مدير دائرة اللوازم والمشتريات هاتف رقم (+97092389687) داخلي (6400) أو من خلال البريد الالكتروني tender3@najah.edu .

مدير دائرة اللوازم والمشتريات

إياد المكاوي



تعديل كراسة طرح عطاء رقم – RFQ-H-11781


جدول الأسعار والكميات :

No.	NNUH#	Item	QTY	Unit	Code + Brand name	Unit Price (Nis)	Total Price (Nis)
1	G-A-MDD00014	Ventilators حسب المواصفات المرفقة في الورقة 6+5+4+3	4	Pcs			
-	Total Price (Nis)		-	-			

lm

	إسم الشركة	1
	الختم	2
	التوقيع والإسم	3
	معلومات الإتصال	4
	مدة التوريد	5
	مدة الكفالة إن وجدت	6

المواصفات الفنية كما في المرفقات أدناه :





ICU Ventilator, Qty: 4	Compliance Yes/No	Deviation
Good brand name with FDA or CE preferable (Europe, USA, Japan)		
Should have CB test and EMC test (60601.1 and 60601.1.2)		
Should have ISO 13485 certification for medical devices		
Manufacturer:		
Source and origin:		
Model:		
ICU Ventilator Specification:		
State of the art technology, the latest version, and the best technology from the manufacturer		
Advanced technology ventilator for use in ICU		
ICU ventilator on trolley		
Suitable for ventilating patients from pediatric to adults		
Should have facility for Invasive and Non-Invasive ventilation works with all modes of ventilation		
Integrated with high flow nasal cannula mode		
Should have permanent O2 sensor		
Please specify the type of flow sensor (Permanent, Reusable or Disposable)		
Standard hinged arm holder for holding the circuit		
Modular design, easy for operation and maintenance		
It should have back up battery with operating time not less than 2 h		
Autoclavable inspiration and expiration assembly		
Audible and visual alarms		
Ventilator with Nebulizer		
Touch screen of minimum of 12"		
Tidal volume: (Minimum at least 50ml, Maximum up to 2000ml) or better		
Flow (in HiFlowO2) (l/min): 2 to 60 or better		
Pressure & flow Trigger		
Flow trigger (l/min): 0.5 to 20 or better		
Pressure trigger (cmH2O) -0.1 to -15.0 or better		
O2 boost function		
Oxygen (%): 21 to 100		



I:E ratio: (1:9 to 4:1) or better		
Respiratory Rate (b/min): 1 to 80		
Apnea back up: On/Off		
Automatic tube compensation		
Leakage compensation		
PEEP/CPAP (cmH2O): 0 to 50 or better		
Dynamic lung display		
Pressure support (cmH2O): 0 to 100 or better		
Tpause (s): 0 to 30		
Trending of measured and calculated values, Trend time 1 to 72 hours.		
Manual ventilation		
Stress Index		
Nebulizer Duration (min): 5 to 40, continuous or better		
Inspiratory and expiratory hold		
Inspiratory flow up to 150 Lpm.		
Real time waveforms: 1- Airway pressure 2- Flow 3- Volume		
Ventilation Modes:		
1) Invasive ventilation – modes:		
Volume Control (VC)		
Pressure Control (PC)		
Pressure Regulated Volume Control (PRVC)		
Pressure Support (PS)		
Volume Support (VS)		
PC-SIMV, VC-SIMV and PRVC-SIMV with pressure support		
BIPAP, CPAP		
2) Non invasive ventilation – modes:		
Pressure Control (PC)		
Pressure Support (PS)		
BIPAP, CPAP		
HFNC		
Apnea Ventilation		
Complete with all accessories and consumables needed to work completely as specified		



Training included:		
Operation training for the Hospital nursing team must be provided on-site		
Service training for one biomedical engineer in the manufacturer site		
Complete with the following accessories:		
2X Expiratory assembly		
O2 and Air hoses with british adapter		
All Accessories for Nebulizer		
1X Adult test Lung		
All HFNC accessories for 5 patients		
Holder for Humidifier		
Optional to be priced separately:		
1- List price all accessories necessary to operate the solution as per the required specs listed above		
2- List all other accessories and supplies that you may recommend to NNUH and are considered optional items to be decided on purchasing at the discretion of NNUH		
3- Battery		
4- Nebulizer kit		
5- CO2 Capnography, mainstream		
6- Complete humidifier with adult and pediatric chamber		
7- Flow sensor		
8- Annual PM kit		

Your offer should contain the following:		
1) Preventive maintenance schedule		
2) Reference list for the same model, the supplier must have 3 reference at least on local market or regional, and must prepare site visit for our team to evaluate the machine.		
3) Compliance sheet		
4) Catalogues		
5) Data sheet		
6) Authorization letter		
7) CE or FDA and ISO certificates and ISO 13485 certificates		
8) CB test and EMC test certificate		



9) The service and operational manuals must delivered as soft copy with the units

10) Guarantee for 10 years for spare parts supply

Warranty: 3 years, including the calibration, software upgrade, with commitment to conduct preventive maintenance (labor and parts) during the warranty period and according to manufacturer recommendations.

Preventive maintenance should be done as per the manufacturer's recommendations, all parts including PM kits to be replaced during the warranty period will be the responsibility of the supplier to provide on his own account. A schedule of the PM visits should be submitted to the hospital biomedical engineering dep.

Hardware (service) keys, software passwords, website account or any means needed to grant the hospital engineer full access to all service functionalities are the right of the hospital and should be delivered with the equipment, and should as well be valid throughout the whole lifetime of machine.



طرح عطاء

يعلن مستشفى النجاح الوطني الجامعي عن طرح عطاء رقم RFQ-H-11781 والخاص بأجهزة Ventilators ، على الشركات المتخصصة والراغبة بالمشاركة بالعطاء تقديم عروض أسعارها في موعد أقصاه يوم الأحد الموافق 2025/02/16 الساعة الثالثة عصراً 03:00 pm .

وقت الإغلاق		موضوع الطرح	نصن نسخة العطاء	موضوع العطاء	رقم العطاء	#
الساعة	التاريخ					
P.M 15:00	2025/02/16	2025/02/04	نصن نسخة 250 شيكال غير مستردة	Ventilators	RFQ-H-11781	1

حسب الشروط التالية:-

1. ضرورة إرفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر .
2. رسوم الإعلان على من يرسو عليه العطاء .
3. ثمن نسخة العطاء 250 شيكل غير مستردة .
4. كفالة دخول عطاء بقيمة (1000) دولار أمريكي .

على الراغبين بالحصول على نسخة العطاء مراجعة دائرة اللوازم والمشتريات مكتب رقم 6400 / مبنى الإدارة أو عبر الإيميل Tender3@najah.edu وذلك خلال ساعات اللوام الرسمية .

للاستفسار يرجى الاتصال على :-

لأية استفسارات أخرى يرجى مراجعة دائرة اللوازم والمشتريات هاتف رقم 2331471 -970-09 +970-9-2389685 فاكس : 6400 / 970-9-2389685 .