



التاريخ: 14/03/2022

طرح عطاء رقم RFQ-H-11483

السادة / الشركات المختصة الراغبة بالمشاركة بالعطاء - المحترمين

الموضوع : **Anesthesia machine for MRI** , حسب المواصفات .

بالإشارة إلى الموضوع أعلاه، يرجى الاطلاع على المواصفات المرفقة مع مراعاة الشروط التالية :-

1. تكتب الاسعار بالشيك الإسرائيلى وبشكل واضح في المكان المخصص حسب الجدول المرفق وتشمل قيمة الضريبة المضافة، في حال وجود تعارض بين سعر الوحدة ومجموع سعر الوحدة المشتبه، فإن سعر الوحدة هو الذي سيعتمد.
2. يتم تقديم شهادة خصم مصدر في حالة تجاوز المبلغ قيمة 2500 شيكل او ما يعادلها.
3. غرامة تأخير 2% عن كل أسبوع وتبدأ من اليوم الذي يلي موعد التسليم المحدد من قبلكم، و يحق للمستشفى إلغاء أمر الشراء والاحالة على مورد آخر دون تحمل اية مسؤولية تذكر في حالة تأخير التوريد ما يزيد عن اسبوعين .
4. تحديد فترة التوريد في عرض سعركم و عدم ذلك يعني أن التوريد فوري وايضاً أن تكون صلاحية العرض المقدم من طرفكم مدة لا تقل عن ثلاثة أشهر .
5. وسلم عروض الاسعار عبر الايميل nnuh.tender@najah.edu فقط ولن يتم قبول أي عرض سعر الا من خلاله ، على أن يكون عرض السعر يحوي فصلاً بين العرض المالي والعرض الفني أي أن العرض المالي منفصل عن العرض الفني .
6. ضرورة الرد على العطاء سلباً أو ايجاباً.
7. الدفع بعد 90 يوم من تاريخ الإسلام والفحص ومطابقة المواصفات الفنية المطلوبة.
8. يحق للمستشفى تجزئة العطاء وهي غير ملزمة بأقل الأسعار.
9. يحق للمستشفى إلغاء العطاء دون اداء السبب ودون تحمل اية تكاليف مع إعادة رسوم نسخة العطاء للمشاركين في العطاء ان وجدت، لا تتحمل المستشفى تكاليف النقل والتحميل والتغذيل.
10. ضرورة ارفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر .
11. اخر موعد لتقديم عرض السعر هو يوم الاثنين الموافق 21/03/2022 الساعة الثالثة مساءاً (15:00PM).
12. لأية استفسارات متعلقة بالعطاء المذكور أعلاه ، يرجى الاتصال بالسيد مدير دائرة اللوازم والمشتريات هاتف رقم tender3@najah.edu (6400) داخلي أو من خلال البريد الإلكتروني 0097092389687 فقط.

مع فائق الاحترام ،،،،

مدير دائرة اللوازم والمشتريات

إياد مكاوي





التاريخ: 14/03/2022م

طرح عطاء رقم RFQ-H-11483

No	Item Code	Item	Unit	QTY	CODE NO. - وكيل حصري / موزع	Unit Price (Nis)	Total Price (Nis)
1	G-A-MDD00170	Anesthesia machine for MRI كما في المواصفات المرفقة في الجدول أدناه	1	PCS			
...	...	Total Price (Nis)	

Anesthesia machine for MRI, Qty: 1	Compliance	Deviation
Good brand name with FDA or CE		
Manufacturer:		
Source and origin:		
Model:		
Anesthesia machine for MRI Specification:		
MRI compatible anesthesia machine		
All Patient type: neonatal, pediatric and adult		
Latest safety alarms as applicable to MRI room and gas failure alarms.		
Features facilitate use in both the MRI suite and surgical OR		
Anesthesia machine with		
- Two Drawers		
- Top shelf and work area		
- Extra shelf for patient monitor		
- Built in ventilator		
- Central wheels lock		
- Ventilator screen		
- Absorber bag arm		
- Canister		
- Two vaporizer (Isoflurane and servoflurane)		
- CO2 absorber		
- Flowmeters for Oxygen, air and nitrous oxide		
- Battery backup		
- Anesthetic gas scavenging system		
- Patient circuit and test lugs for adult and pediatrics.		
The Machine should have a built-in anesthesia ventilator		



Should be compact, ergonomic & easy to use with automatic pre-use check for electronic parts.		
Central brake efectively enables fast and easy interlock or release for enhanced safety especially in magnetic fields		
Should be able to deliver a tidal volume from 20ml to 1500ml.		
Should have a battery backup for 60 minutes with low battery alarm and over charge protection.		
Should have monitoring facility of continuous airway pressure, tidal volume, frequency , oxygen concentration and oxygen supply pressure		
High resolution TFT color display (7")		
Should have automatic self test for the entire system.		
Anesthesia machine should be with 3 gas supply system (O2, N2O and Air) with pipeline connections and reserve cylinder yokes.		
Gas cylinder (pin indexed) yokes		
One Pin index yoke for connecting cylinder each for O2, N2O through pipeline		
Should have pressure gauge for all gas inlets		
Oxygen and Nitrous oxide should be linked either mechanically or pneumatically to ensure a minimum of 25% oxygen delivery at all times to avoid delivery of hypoxic mixture.		
Having mechanical hypoxic guard with automatic cutoff of N2O		
Should provide with oxygen flush switch		
Breathing Frequency (rate) 4 to 60 bpm		
Positive End Expiratory Pressure (PEEP) 0 to 20 cmH2O (hPa)		
Inspiration/expiration ratio (Ti:Te) 4:1 to 1:4		
Ventilation Modes:		
Man/Spon and CPAP		
Volume Controlled (VC)		
Pressure Controlled (PC)		
Pressure Support		
APRV		
SIMV		
The offer includes the following:		
Should provide with adult and pediatric reusable and autoclavable tubing breathing circuit, Qty.1 each		
Isoflurane & servoflurane vaporizer, Qty.1 each		
O2, Air and N2O hose with british standard		
Two vaporizer (Isoflurane and servoflurane) with its filling accessories		
Service training for one hospital engineer in the manufacturer site		



Optional to be priced seperately

System battery pack

O2 cell

PM kit

Gas analyzer

Your offer should contain the following:

- 1) Preventive maintenance schedule
- 2) Reference list for the same model
- 3) Compliance sheet
- 4) Catalogues
- 5) Data sheet
- 6) Authorization letter
- 7) CE or FDA certificate
- 8) The service and operational manuals should be delivered as soft copy with the units

Warranty: 3 years, with commitment to conduct preventive maintenance (labor and parts) during the warranty period and according to manufacturer recommendations.

- يرجى من حضرتكم تحديد مدة التوريد
ضرورة الالتزام بالتسuir على النموذج المرفق بالعطاء، والالتزام بالعملة المحددة(شيك).

- يرجى من حضرتكم الالتزام بتبني المعلومات في الجدول أدناه:

	اسم الشركة	1
	الختم	2
	التوقيع والاسم	3
	معلومات الاتصال	4
	مدة التوريد	5
	مدة الكفالة إن وجدت	6