

التاريخ : 20/05/2025 م

طرح عطاء رقم – RFQ-H-11819

السادة / الشركات المختصة الراغبة بالمشاركة بالعطاء - المحترمين .

الموضوع : **Electrical Bed - Trade In** ، حسب المواصفات المرفقة .

بالإشارة إلى الموضوع أعلاه، يرجى الاطلاع على جداول المواصفات المرفقة مع مراعاة الشروط التالية :-

1. تكتب الاسعار بالشيكال الاسرائيلي وبشكل واضح في المكان المخصص حسب الجدول المرفق وتشمل قيمة الضريبة المضافة، في حال وجود تعارض بين سعر الوحدة ومجموع سعر الوحدة المشتق، فإن سعر الوحدة هو الذي سيعتمد .
2. ثمن نسخة العطاء (250) شيكل إسرائيلي غير مستردة تدفع في الدائرة المالية مبنى الإدارة - الطابق الأول .
3. كفالة دخول عطاء بقيمة (1000) ألف دولار امريكي او ما يعادلها سارية المفعول لمدة ثلاثة شهور تأميناً نقدياً او كفالة بنكية او شيكاً مصدقاً وفي حالة عدم تقديم المطلوب لا ينظر في عرض سعركم .
4. كفالة حسن تنفيذ بقيمة 10% من قيمة العطاء في حالة الترسية تأميناً نقدياً او كفالة بنكية او شيكاً مصدقاً سارية المفعول لمدة ثلاثة اشهر .
5. رسوم الاعلان على من يرسو عليه العطاء .
6. ضرورة الالتزام بالتسعير على النموذج المرفق بالعطاء مع ضرورة تعبئة جدول المواصفات الفنية.
7. أن تكون الأجهزة المعروضة جديدة وليست مجددة .
8. يتم تقديم شهادة خصم مصدر في حالة تجاوز المبلغ قيمة 2500 شيكل او ما يعادلها .
9. غرامة تأخير 2% عن كل اسبوع وتبدأ من اليوم الذي يلي موعد التسليم المحدد من قبلكم، وفي حال تأخير التوريد مدة تزيد عن اسبوعين يحق للمستشفى إلغاء أمر الشراء والاحالة على مورد آخر مع تحميلكم فروق الأسعار .
10. الدفع خلال 90 يوماً من تاريخ الاستلام والفحص ومطابقة المواصفات الفنية المطلوبة .
11. تحديد فترة التسليم في عرض سعركم وايضاً فترة الضمان وأن تكون صلاحية العرض المقدم من طرفكم مدة لا تقل عن ثلاثة اشهر .
12. تسلم عروض الاسعار في مكتب مدير دائرة اللوازم والمشتريات رقم (6400) مبنى الإدارة في مستشفى النجاح الوطني الجامعي بالظرف المختوم فقط وتوضع في صندوق العطاءات ، على أن يكون عرض السعر يحوي فصلاً بين العرض المالي والعرض الفني أي أن العرض المالي منفصل عن العرض الفني مع ضرورة كتابة رقم العطاء على المغلفين وتوضيح اي منهما العرض الفني والعرض المالي .
13. على كل من يرغب بالمشاركة بتقديم عرض سعر تقديم ما لديه من مواد او اجهزة او ماركات لما جاء اعلاه او البديل له والرد على العطاء سلباً او ايجاباً .
14. يحق للمستشفى تجزئة العطاء وهي غير ملزمة بأقل الأسعار .
15. يحق للمستشفى إلغاء العطاء دون ابداء السبب ودون تحمل اية تكاليف مع اعادة رسوم نسخة العطاء للمشاركين في العطاء ان وجدت، ولا تتحمل المستشفى تكاليف النقل والتحميل والتنزيل .
16. ضرورة ارفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر .
17. آخر موعد لتقديم عرض السعر هو يوم الأحد الموافق 2025-06-01 الساعة الواحدة ظهراً PM 01:00 .
18. لأية إستفسارات متعلقة بالعطاء المذكور اعلاه ، يرجى الاتصال بالسيد مدير دائرة اللوازم والمشتريات هاتف رقم (+97092389687) داخلي (6400) أو من خلال البريد الالكتروني tender3@najah.edu .

مدير دائرة اللوازم والمشتريات

إياد المكاوي



التاريخ : 20/05/2025 م


طرح عطاء رقم – RFQ-H-11819**جدول الأسعار والكميات :**

No	NNUH Code	Item	QTY	Unit	CODE #	وكيل / موزع	مدة التوريد	U.P (NIS)	T.P (NIS)
1	G-A-MDQ01225	Trade In Electrical Bed حسب المواصفات الفنية	30	Pcs					
-		Total Price (NIS)							

يرجى تعبئة الجدول التالي مع ضرورة الختم على جميع أوراق عرض السعر.

1	إسم الشركة
2	الختم
3	التوقيع والإسم
4	معلومات الإتصال
5	مدة التوريد
6	مدة الكفالة إن وجدت

المواصفات الفنية :



Electrical Patient Bed, Qty: 30	Compliance Yes/No	Deviation
Good brand name with FDA or CE preferable (Europe, USA, Japan)		
Should has CB test and EMC test (60601.1 and 60601.2)		
Should has ISO 13485 certification for medical device		
Manufacturer:		
Source and origin:		
Model:		
Please make trade in offer with our stieglmeyer mechanical bed (DEKA), Qty. 30		
Electrical Patient Bed Specification:		
Heavy duty electrical hospital bed shall be suitable for 24/7 clinical use in busy ward settings		
Must be designed for adult patient care and meet relevant hospital safety and comfort standards		
The bed must maintain stability during all movement operations (raising/lowering, Trendelenburg, etc.) without tilting, rocking, or skidding. Braking systems must lock the bed completely with minimal foot pressure		
Electrically operated and controlled		
4 motors electric bed		
With fixed headboard which does not move during height adjustments, prevent collisions with furniture during height adjustment		
Side rails, headboard and footboard panels are made from polypropylene materials		
Safe working load approximately 250 kg		
The mattress base width approximately 90 cm		
The mattress base length approximately 200 cm		
Can be accommodated in an elevator with 125 cm by 275 cm dimensions		
4 sections mattress base, and should be with polypropylene cover, easy to clean, mattress base elements can be removed for reprocessing		
Central brake		
Cardiac chair and fowler positions		
With nurse control and two double-face siderail control units		
The nurse control is housed in the bed linen holder and they can attached to the footboard		



All cables integrated in the bed frame		
CPR liver and/or electrical CPR button		
Manual CPR function for backrest		
Bed frame constructed with high-strength powder-coated steel, non rusty material frame		
Easy cleaning concept, no inaccessible recesses in the headboard or footboard and safety sides		
With corner bumpers to protect the bed		
Sterilizable mattress		
Waterproof, antimicrobial hospital-grade mattress included		
High protection against water and dust, please specify the IP rating? Higher is better		
Built in angle indecator (at the both side of the bed) to detect the trendelenburg angle		
Built in angle indecator (at the both side of the bed) to detect the back section angle		
The following movements must be electrically available with the following bed angles: (Controlled via siderail control and nurse control panel)		
a- Backrest (Head section): 0 to 70 degree approximately		
b- Knee rest (leg section): 0 to 40 degree approximately		
c- Trendelenburg and Reverse Trendelenburg: (-15 to +15) degree approximately		
d- Adjustable height: (35-80) cm approximately		
Urinary catheter holder (under the bed)		
Compatible with universal disinfectant agents		
AC input: 220V, 50Hz		
Complete with all accessories and consumables needed to work completely as specified		
Training included:		
Operation training for the Hospital nursing team must be provided on-site		
Service training for hospital engineers must be provided on-site		
The machine should be supplied with the following Accessories:		
1- IV pole		
2- Mattress		
3- Urinary catheter holder		
4- Main nurse control unit		
5- Side rail control unit		

Optional to be priced separately:		
1- List price all accessories necessary to operate the solution as per the required specs listed above		
2- List all other accessories and supplies that you may recommend to NNUH and are considered optional items to be decided on purchasing at the discretion of NNUH		
3- Main control box		
4- Main nurse control unit		
5- Actuator		
6- Side rail control unit		
7- Castor		

Your offer should contain the following:		
1) Preventive maintenance schedule		
2) Reference list for the same model, the supplier must have 3 reference at least on local market or regional, and must prepare site visit for our team to evaluate the product.		
3) Compliance sheet		
4) Catalogues		
5) Data sheet		
6) Authorization letter		
7) CE or FDA and ISO certificates		
8) CB test and EMC test certificate		
9) The service and operational manuals must delivered as soft and hard copy with the units		
10) Guarantee for 10 years for spare parts supply		

Warranty: 3 years, including the calibration, software upgrade, with commitment to conduct preventive maintenance (labor and parts) during the warranty period and according to manufacturer recommendations.		
Preventive maintenance should be done as per the manufacturer's recommendations, all parts including PM kits to be replaced during the warranty period will be the responsibility of the supplier to provide on his own account. A schedule of the PM visits should be submitted to the hospital biomedical engineering dep.		
Hardware (service) keys, software passwords, website account or any means needed to grant the hospital engineer full access to all service functionalities are the right of the hospital and should be delivered with the equipment, and should as well be valid throughout the whole lifetime of machine.		



طرح عطاء

يعلن مستشفى النجاح الوطني الجامعي عن طرح عطاء رقم RFQ-H-11819 والخاص ب Trade in – Electrical Bed ، على الشركات المتخصصة والراغبة بالمشاركة بالعطاء تقديم عروض اسعارها في موعد اقصاه يوم الخميس الموافق 2025/04/17 م الساعة الواحدة ظهراً PM 13:00 :

#	رقم العطاء	موضوع العطاء	ثمن نسخة العطاء	موعد الطرح	وقت الإغلاق	
					التاريخ	الساعة
1	RFQ-H-11819	Electrical Bed – Trade in	250 شيكل غير مستردة	2025/05/20	2025/06/01	P.M 13 :00

ب الشروط التالية:

1. ضرورة ارفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر .
2. رسوم الاعلان على من يرسو عليه العطاء .
3. ثمن نسخة العطاء 250 شيكل غير مستردة .
4. كغالة دخول عطاء بقيمة (1000) دولار أمريكي .

الراغبين بالحصول على نسخة العطاء مراجعة دائرة اللوازم والمشتريات مكتب رقم 6400 / مبنى الإدارة أو عبر الإيميل Tender3@najah.edu وذلك خلال ساعات الدوام الرسمية .

بمقتضى يرجى الاتصال على :

استفسارات أخرى يرجى مراجعة دائرة اللوازم والمشتريات هاتف رقم 2331471-970-09+ تحويلة 6400 / فاكس: 2389685-970-9+ .