



التاريخ : 2024/12/31 م

طرح عطاء رقم RFQ-H-11767

السادة / الشركات الراغبة بالمشاركة بالعطاء - المحترمين .

الموضوع : ABR Machine ، حسب المواصفات.

بالإشارة إلى الموضوع أعلاه، يرجى الاطلاع على المواصفات المرفقة مع مراعاة الشروط التالية :-

1. تكتب الاسعار بالشيكول الإسرائيلي وبشكل واضح في المكان المخصص حسب الجدول المرفق و تشمل قيمة الضريبة المضافة, في حال وجود تعارض بين سعر الوحدة ومجموع سعر الوحدة المشتق، فإن سعر الوحدة هو الذي سيعتمد .
2. يتم تقديم شهادة خصم مصدر في حالة تجاوز المبلغ قيمة 2500 شيكل او ما يعادلها.
3. غرامة تأخير 2% عن كل اسبوع وتبدأ من اليوم الذي يلي موعد التسليم المحدد من قبلكم , و يحق للمستشفى إلغاء أمر الشراء والاحالة على مورد آخر دون تحمل اية مسؤولية تذكر في حالة تأخير التوريد ما يزيد عن اسبوعين .
4. تحديد فترة التوريد في عرض سعركم **وعدم ذلك يعنى أن التوريد فوري** وايضاً أن تكون صلاحية العرض المقدم من طرفكم مدة لا تقل عن ثلاثة أشهر .
5. يجب أن لا تقل صلاحية الأصناف المسعرة عن عامين من تاريخ التوريد.
6. **تسلم عروض الاسعار بالظرف المغلق فقط** ، على ان يكون عرض السعر يحوي فصلاً بين العرض المالي والعرض الفني اي ان العرض المالي منفصل عن العرض الفني .
7. ضرورة الرد على العطاء سلباً أو ايجاباً.
8. ضرورة الالتزام بالتسعير على جدول المواصفات الكميات في نسخة طرح العطاء.
9. الدفع خلال 90 يوم من تاريخ الإستلام والفحص ومطابقة المواصفات الفنية المطلوبة.
10. يحق للمستشفى تجزئة العطاء وهي غير ملزمة بأقل الأسعار.
11. يحق للمستشفى إلغاء العطاء دون ابداء السبب ودون تحمل اية تكاليف مع إعادة رسوم نسخة العطاء للمشاركين في العطاء ان وجدت، لا تتحمل المستشفى تكاليف النقل والتحميل والتتزيل.
12. ضرورة ارفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر .
13. **اخر موعد لتقديم عرض السعر هو يوم الخميس الموافق 2025/01/09 الساعة العاشرة صباحاً (10:00) AM.**
14. لأية إستفسارات متعلقة بالعطاء المذكور اعلاه , يرجى الاتصال بالسيد مدير دائرة اللوازم والمشتريات هاتف رقم (0097092389687) داخلي (6400) أو من خلال البريد الالكتروني tender3@najah.edu .

مع فائق الاحترام ,,,,

مدير دائرة اللوازم والمشتريات

إياد مكاوي

التاريخ : 2024/12/31 م

طرح عطاء رقم RFQ-H-11767

جدول المواصفات والكميات

N o.	Item Code	Item	Unit	QTY	Code No	U.P (Nis)	T.P (Nis)
1	G-A-MDD01838	ABR machine حسب المواصفات المرفقة	Pcs	1			
-	Total Price (Nis)						

ABR system, Qty: 1	Compliance	Deviation
Good brand name with FDA or CE preferable (Europe, USA, Japan)		
Should have CB test and EMC test (60601.1 and 60601.2)		
Should has ISO 13485 certification for medical devices		
Manufacturer:		
Source and origin:		
Model:		
Trade in with our old ABR system / GSI / SN: GS0048659		
ABR system Specification:		
State of the art technology and the latest version from the manufacturer and the latest and best technology on the market		
Portable equipment		
ABR – diagnostic and threshold estimation		
Evoked Potential (EP) and Otoacoustic Emissions (OAE) device		
Stimulation: Air-conduction (AC), Bone-conduction (BC) Stimuli: Click, Tone burst, Broadband chirp.		
Performing all standard evoked potential tests: ASSR (auditory steady state responses) and TEOAE (Otoacoustic emissions).		
TEST TYPES • ABR, ECochG, MLR, LLR , SN10, P300/MMN, eABR, oVEMP, cVEMP • ASSR: Binaural Testing, 4 Frequency Simultaneous Testing per ear • DPOAE, TEOAE, Spontaneous OAE		



• Test Stimuli: CE-Chirp, CE-Chirp Octave Bands, Speech Stimuli, Click, Tone Bursts

Modern waveform analysis tools calculation

Automatic SNR (Signal to Noise Ratio) and Residual Noise calculation

Artifact Rejections: Adjustable level (0-100%) and any region within the analysis time window

Adjustable digital filters: High-pass, Low-pass

Measured variables: Real-time Wave: I, II, III, IV, V latencies

Interpeak intervals: I-III, III-V, I-V

Amplitudes: Wave I & V, V/I amplitude ratio Latency-specific

Correlation Coefficient Latency norms: Newborn to adults Masking: White noise, 0-90 dB HL ABR

Screening – automated screening Stimulation: Air-conduction (AC) Stimuli: 30 or 35 dB nHL click 100 μ s ABR detection: Automated Report: Pass / Refer.

Internal database: up to 1000 patient

Complete with all accessories and consumables needed to work completely as specified

Training included in the price:

On site user training

Complete with the following accessories:

1- Suitable computer

Optional to be priced separately:

List price all accessories necessary to operate the solution as per the required specs listed above

List all other accessories and supplies that you may recommend to NNUH and are considered optional items to be decided on purchasing at the discretion of NNUH

Your offer should contain the following:

1) Preventive maintenance schedule

2) Reference list for the same model, the supplier must have 3 reference at least on local market or regional, and must prepare site visit for our team to evaluate the machine.



- | | | |
|---|--|--|
| 3) Compliance sheet | | |
| 4) Catalogues | | |
| 5) Data sheet | | |
| 6) Authorization letter | | |
| 7) CE or FDA and ISO certificates and ISO 13485 certificates | | |
| 8) CB test and EMC test certificate | | |
| 9) The service and operational manuals must delivered as soft copy with the units | | |
| 10) Guarantee for 10 years for spare parts supply | | |

Warranty: 3 years, including the calibration, software upgrade, with commitment to conduct preventive maintenance (labor and parts) during the warranty period and according to manufacturer recommendations.

Preventive maintenance should be done as per the manufacturer's recommendations, all parts including PM kits to be replaced during the warranty period will be the responsibility of the supplier to provide on his own account. A schedule of the PM visits should be submitted to the hospital biomedical engineering dep.

Hardware (service) keys, software passwords, website account or any means needed to grant the hospital engineer full access to all service functionalities are the right of the hospital and should be delivered with the equipment, and should as well be valid throughout the whole lifetime of machine.

- ضرورة الالتزام بالتسعير على النموذج المرفق والتسعير بعملة الشيك فقط .

- يرجى من حضرتكم الالتزام بتعبئة المعلومات في الجدول أدناه:

1	إسم الشركة والختم	
3	الإسم والتوقيع	
4	معلومات الإتصال	
5	مدة التوريد	
6	مدة الكفالة إن وجدت	

An-Najah National University Hospital
Procurement Department



NNUH

مستشفى النجاح الوطني الجامعي
دائرة اللوازم والمشتريات

إعلان طرح عطاء

يعن مستشفى النجاح الوطني الجامعي عن طرح عطاء رقم RFQ-H-11767، على الشركات المتخصصة والراغبة بالمشاركة بالعطاء تقديم عروض اسعرها في موعد أقصاه يوم الخميس الموافق 2025/01/09 الساعة العاشرة صباحاً AM 10:00 وفقاً للآتي:

#	رقم العطاء	موضوع العطاء	ثمن نسخة العطاء	موعد الطرح	وقت الإغلاق	
					التاريخ	الساعة
1	RFQ-H-11767	ABR machine	مجاً	2024/12/31	2025/01/09	AM 10:00

حسب الشروط التالية:

1. ضرورة إرفاق الرخصة التجارية والسيرة الذاتية للشركة مع عرض السعر.

على الراغبين بالاشتراك في العطاء المذكور طلبه بواسطة ارسال رسالة على الإيميل التالي tender3@najah.edu أو مراجعة دائرة اللوازم والمشتريات مكتب رقم 6400 \ في مستشفى النجاح الوطني الجامعي، مبنى الإدارة، للحصول على نسخة من وثائق العطاء

للاستفسال يرجى الاتصال على:

دائرة اللوازم والمشتريات في مستشفى النجاح الوطني الجامعي - البريد الإلكتروني: tender3@najah.edu هاتف: +970-09-2331471 فاكس: +970-9-2389685.

Palestine – Nablus – Asera St.
<https://www.nnuh.org>

Email: tender3@najah.edu

Fax: +970 9 2389685
P.O BOX: 797